

## FFPE DNA Genome In A Bottle Reference Standard

PRODUCT INFORMATION	
Catalog ID	GM24149
Batch ID	41953
Product Description	Ashkenazim PGP Father Reference Standard
Date of Manufacture	28-Nov-2018
Manufacturing QMS Status	ISO 13485

PROPERTIES	
Format	15 µm section
Storage Conditions	+2-8°C
Use by	For the latest information on product shelf life, please visit our website: <a href="https://www.horizondiscovery.com/reference-standards/coa-information">https://www.horizondiscovery.com/reference-standards/coa-information</a>

CUSTOMER SUPPORT	
Technical Support	technical@horizondiscovery.com
Customer Service	orders@horizondiscovery.com

ADDITIONAL INFORMATION	
N/A	

**QUALITY CONTROL**Quality: **Agarose gel electrophoresis**

Acceptance Criteria	Band Present/Absent	Pass/Fail
Bright band of high molecular weight	Present	PASS

Quantification: **QuantiFluor® assay (Promega)\***

Acceptance Criteria	Reading	Pass/Fail
≥ 400 ng (Average yield)	692.56 ng	PASS

\*DNA extraction performed using Maxwell® 16 FFPE plus LEV DNA Purification Kit (Promega)

**FOR RESEARCH USE ONLY**

Manufactured in the UK

**Horizon Discovery**8100 Cambridge Research Park,  
Waterbeach, Cambridge,  
CB25 9TL, United Kingdom  
[www.horizondiscovery.com/](http://www.horizondiscovery.com/)t + 44 (0) 1223 976 000  
f + 44 (0) 1223 655 581  
e [orders@horizondiscovery.com](mailto:orders@horizondiscovery.com)