

FFPE DNA Reference Material

IMPORTANT: Centrifuge tube prior to opening to prevent loss of product

PRODUCT INFORMATION	
Catalogue Number	HD-C1639
Batch ID	
Product Description	Idylla MSI FFPE Reference Standard
Date of Manufacture	
Manufacturing QMS Status	ISO 13485

PROPERTIES	
Format	10 µm section
Storage Conditions	+4°C (2-8°C)
Cell Density	100 x10 ⁶ cells per 1cm block
Use by	For the latest information on product shelf life, please visit our website: https://horizondiscovery.com/products/diagnostic-reference-standards/reference-standards-product-information

CUSTOMER SUPPORT	
Technical	technical@horizondiscovery.com
Orders	orders@horizondiscovery.com

ADDITIONAL INFORMATION

--

QUALITY CONTROL

% of HCT116 cell line: Droplet DigitalTM PCR

Expected % of HCT116 cell line	Measured % of HCT116 cell line	Acceptance Criteria	Pass/Fail
16.67%		13.34 - 20.00%	

Quality: Genomic DNA Screen Tape assay (Agilent Technologies)

Acceptance Criteria	DIN value	Pass/Fail
DIN value >7.0		

FOR RESEARCH USE ONLY

Manufactured in the UK

Horizon Discovery

8100 Cambridge Research Park,
Waterbeach, Cambridge,
CB25 9TL, United Kingdom
www.horizondiscovery.com/

t + 44 (0) 1223 976 000

f + 44 (0) 1223 655 581

e

orders@horizondiscovery.com

03-16-071 Rev 01