

## Singleplex gDNA Reference Standard

**IMPORTANT:** Centrifuge tube prior to opening to prevent loss of product

### PRODUCT INFORMATION

Catalog ID	HD688
Batch ID	35332
Product Description	50% PIK3CA E542K
Date of Manufacture	04-Aug-2017
Manufacturing QMS Status	ISO 13485

### PROPERTIES

Quantity	1 µg
Concentration	50 ng/µL
Buffer	Tris-EDTA (10 mM Tris-HCl, 1 mM EDTA), pH 8.0
Storage Conditions	+4°C
Use by	For the latest information on product shelf life, please visit our website: <a href="https://www.horizondiscovery.com/reference-standards/coa-information">https://www.horizondiscovery.com/reference-standards/coa-information</a>

### CUSTOMER SUPPORT

Technical Support	<a href="mailto:technical@horizondiscovery.com">technical@horizondiscovery.com</a>
Customer Service	<a href="mailto:orders@horizondiscovery.com">orders@horizondiscovery.com</a>

### ADDITIONAL INFORMATION

N/A

#### Horizon Discovery

8100 Cambridge Research Park,  
Waterbeach, Cambridge,  
CB25 9TL, United Kingdom  
[www.horizondiscovery.com/](http://www.horizondiscovery.com/)

t + 44 (0) 1223 976 000  
f + 44 (0) 1223 655 581  
e [orders@horizondiscovery.com](mailto:orders@horizondiscovery.com)

**QUALITY CONTROL**

 Allele Frequency (AF) and Genotype: **Droplet Digital™ PCR**

Expected Allele Frequency	Acceptance Criteria	Actual AF (%)	Pass/Fail
50.00%	45.00-55.00%	50.40	PASS

 Quality: **Agarose gel electrophoresis**

Acceptance Criteria	Band Present/Absent	Pass/Fail
Bright band of high molecular weight	Present	PASS

 Quantification: **Spectrophotometry (A260)**

Acceptance Criteria	Reading	Pass/Fail
48.00 - 52.00 ng/μL	50.24 ng/μL	PASS

**FOR RESEARCH USE ONLY**

Manufactured in the UK

**Horizon Discovery**

8100 Cambridge Research Park,  
 Waterbeach, Cambridge,  
 CB25 9TL, United Kingdom  
[www.horizondiscovery.com/](http://www.horizondiscovery.com/)

t + 44 (0) 1223 976 000  
 f + 44 (0) 1223 655 581  
 e [orders@horizondiscovery.com](mailto:orders@horizondiscovery.com)