



horizon

CERTIFICATE OF ANALYSIS

FFPE DNA Reference Standard (Sections)

IMPORTANT: Centrifuge tube prior to opening to prevent loss of product

PRODUCT INFORMATION

Catalogue No.	HD789	Serial No.	31292
Date of Manufacture	May 2018	Description	Structural Multiplex

PROPERTIES

Format	15 µm section
Expected DNA Yield	≥ 400 ng *
Storage Conditions	+4°C
Use By	Stable for 48 months from date of manufacture

* DNA extraction performed using Maxwell® 16 FFPE plus LEV DNA Purification Kit (Promega)

QUALITY CONTROL

TEST	TEST METHOD	PASS/FAIL
Allelic Frequency	Droplet Digital™ PCR	Pass ^a
Genotype		
Quality	Agarose Gel Electrophoresis	Pass
Quantification	QuantiFluor® assay (Promega)	Pass

^a This product has been released under concession. MYC-N Assay out of specification. MYC-N copy number should be between 6.8 and 10.2 copies; however, this batch exceeds this limit (10.4 copies). All other quality control criteria has passed.

CUSTOMER SUPPORT

Technical	technical@horizondiscovery.com
Orders	orders@horizondiscovery.com

FOR RESEARCH USE ONLY