

## Multiplex gDNA Reference Standard

**IMPORTANT:** Centrifuge tube prior to opening to prevent loss of product

PRODUCT INFORMATION	
Catalog ID	HD795
Batch ID	36703
Product Description	<a href="#">BRCA Somatic Multiplex I gDNA Reference Standard</a>
Date of Manufacture	27-Jul-2018
Manufacturing QMS Status	ISO 13485

PROPERTIES	
Quantity	1 µg
Concentration	50 ng/µL
Buffer	Tris-EDTA (10 mM Tris-HCl, 1 mM EDTA), pH 8.0
Storage Conditions	+4°C
Use by	For the latest information on product shelf life, please visit our website: <a href="https://www.horizondiscovery.com/reference-standards/coa-information">https://www.horizondiscovery.com/reference-standards/coa-information</a>

CUSTOMER SUPPORT	
Technical Support	<a href="mailto:technical@horizondiscovery.com">technical@horizondiscovery.com</a>
Customer Service	<a href="mailto:orders@horizondiscovery.com">orders@horizondiscovery.com</a>

ADDITIONAL INFORMATION	
N/A	

### Horizon Discovery

8100 Cambridge Research Park,  
Waterbeach, Cambridge,  
CB25 9TL, United Kingdom  
[www.horizondiscovery.com/](http://www.horizondiscovery.com/)

t + 44 (0) 1223 976 000  
f + 44 (0) 1223 655 581  
e [orders@horizondiscovery.com](mailto:orders@horizondiscovery.com)

**QUALITY CONTROL**

Allele Frequency (AF) and Genotype: Droplet Digital™ PCR

Gene	Variant	Expected AF (%)	Actual AF (%)	Acceptance Criteria (%)		Pass/Fail
				Low	High	
BRCA2	V2466A	100.0	100.00	90.00	100.00	PASS
BRIP1	S919P	100.0	99.97	90.00	100.00	PASS
NBN	E185Q	60.0	62.40	54.00	66.00	PASS
BARD1	R378S	42.5	45.71	38.25	46.75	PASS
BRCA2	N1784fs	40.0	39.40	36.00	44.00	PASS
BRCA1	R1443STOP	32.5	31.50	29.25	35.75	PASS
BRCA2	K1691fs	32.5	32.20	29.25	35.75	PASS
BRCA2	D1420Y	32.5	31.70	29.25	35.75	PASS
BRCA1	P871L	15.0	15.95	13.50	16.50	PASS
BRCA2	I2675fs	10.0	10.23	8.00	12.00	PASS
BRCA1	S1613G	7.5	7.87	6.00	9.00	PASS
BRCA1	K1183R	7.5	7.97	6.00	9.00	PASS
BRCA1	K820E	7.5	7.50	6.00	9.00	PASS
BRCA1	D435Y	7.5	8.08	6.00	9.00	PASS
BRCA2	N289H	7.5	7.38	6.00	9.00	PASS
BRCA2	N991D	7.5	7.66	6.00	9.00	PASS

Quality: Agarose gel electrophoresis\*

Acceptance Criteria	Band Present/Absent	Pass/Fail
Bright band of high molecular weight	Present	PASS

Quantification: Spectrophotometry (A260)\*

Acceptance Criteria	50 ng/μL ±4% (Accepted Range: 48-52 ng/μL)	PASS

\*Quality Control Test occurs before gDNA is multiplexed

**FOR RESEARCH USE ONLY**

Manufactured in the UK

**Horizon Discovery**

8100 Cambridge Research Park,  
Waterbeach, Cambridge,  
CB25 9TL, United Kingdom  
www.horizondiscovery.com/

t + 44 (0) 1223 976 000  
f + 44 (0) 1223 655 581  
e orders@horizondiscovery.com