

Multiplex Formalin Compromised DNA Reference Standard

IMPORTANT: Centrifuge tube prior to opening to prevent loss of product

PRODUCT INFORMATION	
Catalog ID	HD799
Batch ID	
Product Description	Quantitative Multiplex Formalin Compromised (Moderate) Reference Standard
Date of Manufacture	
Manufacturing QMS Status	ISO 9001

PROPERTIES	
Quantity	1 µg
Concentration	50 ng/µL
Buffer	Tris-EDTA (10 mM Tris-HCl, 1 mM EDTA), pH 8.0
Storage Conditions	+4°C
Use By	For the latest information on product shelf-life, please visit our website: https://www.horizondiscovery.com/reference-standards/coa-information

CUSTOMER SUPPORT	
Technical Support	technical@horizondiscovery.com
Customer Service	orders@horizondiscovery.com

ADDITIONAL INFORMATION	

QUALITY CONTROL

 Allele Frequency (AF) and Genotype: **Droplet Digital™ PCR**

Gene	Variant	Expected AF (%)	Actual AF (%)	Acceptance Criteria (%)		Pass/Fail
				Low	High	
BRAF	V600E	10.5		7.40	13.70	
cKIT	D816V	10.0		7.00	13.00	
EGFR	ΔE746-A750	2.0		1.20	2.80	
EGFR	L858R	3.0		1.80	4.20	
EGFR	T790M	1.0		0.60	1.40	
EGFR	G719S	24.5		19.60	29.40	
KRAS	G13D	15.0		10.50	19.50	
KRAS	G12D	6.0		4.20	7.80	
NRAS	Q61K	12.5		8.80	16.30	
PIK3CA	H1047R	17.5		12.30	22.80	
PIK3CA	E545K	9.0		6.30	11.70	

 Quantification: **Qubit® dsDNA BR assay (Invitrogen)**

Acceptance Criteria	Reading	Pass/Fail
45-55 ng/μL		

 Fragmentation: **D1000 DNA ScreenTape assay (Agilent Tech)**

Acceptance Criteria	Reading	Pass/Fail
DIN 2.5 - 5.0		

FOR RESEARCH USE ONLY

Manufactured in the UK

Horizon Discovery

 8100 Cambridge Research Park,
 Waterbeach, Cambridge,
 CB25 9TL, United Kingdom
www.horizondiscovery.com/

 t + 44 (0) 1223 976 000
 f + 44 (0) 1223 655 581
 e orders@horizondiscovery.com