

## Multiplex Formalin Compromised DNA Reference Standard

**IMPORTANT:** Centrifuge tube prior to opening to prevent loss of product

PRODUCT INFORMATION	
Catalog ID	HD803
Batch ID	47192
Product Description	<a href="#">Quantitative Multiplex Formalin Compromised (Severe) Reference Standard</a>
Date of Manufacture	11-Feb-2021
Manufacturing QMS Status	ISO 9001

PROPERTIES	
Quantity	1 µg
Concentration	50 ng/µL
Buffer	Tris-EDTA (10 mM Tris-HCl, 1 mM EDTA), pH 8.0
Storage Conditions	+4°C (2-8°C)
Use by	For the latest information on product shelf life, please visit our website: <a href="https://horizondiscovery.com/products/diagnostic-reference-standards/reference-standards-product-information">https://horizondiscovery.com/products/diagnostic-reference-standards/reference-standards-product-information</a>

CUSTOMER SUPPORT	
Technical Support	<a href="mailto:technical@horizondiscovery.com">technical@horizondiscovery.com</a>
Customer Service	<a href="mailto:orders@horizondiscovery.com">orders@horizondiscovery.com</a>

ADDITIONAL INFORMATION	
N/A	

### Horizon Discovery

8100 Cambridge Research Park,  
Waterbeach, Cambridge,  
CB25 9TL, United Kingdom  
[www.horizondiscovery.com/](http://www.horizondiscovery.com/)

t + 44 (0) 1223 976 000  
f + 44 (0) 1223 655 581  
e [orders@horizondiscovery.com](mailto:orders@horizondiscovery.com)

**QUALITY CONTROL**

## Allele Frequency (AF) and Genotype: Droplet Digital™ PCR

Gene	Variant	Expected AF (%)	Actual AF (%)	Acceptance Criteria (%)		Pass/Fail	
				Low	High		
BRAF	V600E	10.5	13.30	7.40	13.70	PASS	▼
KIT	D816V	10.0	9.90	7.00	13.00	PASS	▼
EGFR	ΔE746-A750	2.0	2.06	1.20	2.80	PASS	▼
EGFR	L858R	3.0	3.88	1.80	4.20	PASS	▼
EGFR	T790M	1.0	1.02	0.60	1.40	PASS	▼
EGFR	G719S	24.5	23.10	19.60	29.40	PASS	▼
KRAS	G13D	15.0	15.20	10.50	19.50	PASS	▼
KRAS	G12D	6.0	6.59	4.20	7.80	PASS	▼
NRAS	Q61K	12.5	11.40	8.80	16.30	PASS	▼
PIK3CA	H1047R	17.5	19.80	12.30	22.80	PASS	▼
PIK3CA	E545K	9.0	8.08	6.30	11.70	PASS	▼

## Quantification: Qubit® dsDNA BR assay (Invitrogen)

Acceptance Criteria	Reading	Pass/Fail
45-55 ng/μL	46.87 ng/μL	PASS ▼

## Fragmentation: D1000 DNA ScreenTape assay (Agilent Tech) TapeStation software version A.02.02

Acceptance Criteria	Reading	Pass/Fail
DIN ≤ 2.0	1.9	PASS

**FOR RESEARCH USE ONLY**

Manufactured in the UK

**Horizon Discovery**

 8100 Cambridge Research Park,  
 Waterbeach, Cambridge,  
 CB25 9TL, United Kingdom  
[www.horizondiscovery.com/](http://www.horizondiscovery.com/)

 t + 44 (0) 1223 976 000  
 f + 44 (0) 1223 655 581  
 e [orders@horizondiscovery.com](mailto:orders@horizondiscovery.com)