

Multiplex FFPE DNA Reference Standard

PRODUCT INFORMATION

Catalog ID	HD810
Batch ID	
Product Description	BRCA Somatic Multiplex I FFPE DNA Reference Standard
Date of Manufacture	
Manufacturing QMS Status	ISO 13485

PROPERTIES

Format	15 µm section
Storage Conditions	+4°C (2-8°C)
Use by	For the latest information on product shelf life, please visit our website: https://horizondiscovery.com/en/reference-standards/reference-standards-product-information

CUSTOMER SUPPORT

Technical Support	technical@horizondiscovery.com
Customer Service	orders@horizondiscovery.com

ADDITIONAL INFORMATION

Horizon Discovery

8100 Cambridge Research Park,
Waterbeach, Cambridge,
CB25 9TL, United Kingdom
www.horizondiscovery.com/

t + 44 (0) 1223 976 000
f + 44 (0) 1223 655 581
e orders@horizondiscovery.com

QUALITY CONTROL

Allele Frequency (AF) and Genotype: Droplet Digital™ PCR

Gene	Variant	Expected AF (%)	Actual AF (%)	Acceptance Criteria (%)		Pass/Fail
				Low	High	
BRCA2	V2466A	100.0		90.00	100.00	
BRIP1	S919P	100.0		90.00	100.00	
NBN	E185Q	60.0		54.00	66.00	
BARD1	R378S	42.5		38.30	46.80	
BRCA2	N1784fs	40.0		36.00	44.00	
BRCA1	R1443STOP	32.5		29.30	35.80	
BRCA2	K1691fs	32.5		29.30	35.80	
BRCA2	D1420Y	32.5		29.30	35.80	
BRCA1	P871L	15.0		12.00	18.00	
BRCA2	I2675fs	10.0		8.00	12.00	
BRCA1	S1613G	7.5		6.00	9.00	
BRCA1	K1183R	7.5		6.00	9.00	
BRCA1	K820E	7.5		6.00	9.00	
BRCA1	D435Y	7.5		6.00	9.00	
BRCA2	N289H	7.5		6.00	9.00	
BRCA2	N991D	7.5		6.00	9.00	

Quality: Agarose gel electrophoresis

Acceptance Criteria	Band Present/Absent	Pass/Fail
Bright band of high molecular weight		

Quantification: QuantiFluor® assay (Promega) *

Acceptance Criteria	Reading	Pass/Fail
≥ 400 ng (Average yield)		

*DNA extraction performed using Maxwell® RSC FFPE plus DNA Kit (Promega)

FOR RESEARCH USE ONLY

Manufactured in the UK

Horizon Discovery

8100 Cambridge Research Park,
 Waterbeach, Cambridge,
 CB25 9TL, United Kingdom
www.horizondiscovery.com/

t + 44 (0) 1223 976 000
 f + 44 (0) 1223 655 581
 e orders@horizondiscovery.com